

Concurrent Validity of a Continuous Glucose-Monitoring System at Rest and During and Following a High-Intensity Interval Training Session

Pauline Clavel, Marina Fabre, Cédric Leduc, Eve Tiollier, Mathieu Lacome,

Martin Buchheit

▶ To cite this version:

Pauline Clavel, Marina Fabre, Cédric Leduc, Eve Tiollier, Mathieu Lacome, et al.. Concurrent Validity of a Continuous Glucose-Monitoring System at Rest and During and Following a High-Intensity Interval Training Session. International Journal of Sports Physiology and Performance, 2022, 17 (4), pp.627-633. 10.1123/ijspp.2021-0222. hal-03881248

HAL Id: hal-03881248 https://insep.hal.science//hal-03881248

Submitted on 1 Dec 2022

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

- *Title:* Concurrent validity of a continuous glucose monitoring system at rest, during and
 following a high-intensity interval training session.
- 3
- 4 Submission type: Original research
- 5 *Authors:* Clavel $P^{1,2}$, Tiollier E^2 , Leduc $C^{1,3}$, Fabre $M^{1,2}$, Lacome $M^{1,2}$, Buchheit $M^{2,4,5,6}$
- ⁶ ¹ Performance Department, Paris Saint-Germain FC, Saint-Germain-en-Laye, France
- ⁷ ² French National Institute of Sport (INSEP), Laboratory of Sport, Expertise and Performance
- 8 (EA 7370), Paris, France
- 9 ³ Carnegie Applied Rugby Research (CARR) centre, Institute for Sport, Physical Activity and
- 10 Leisure, Carnegie School of Sport, Leeds Beckett University, Leeds, United Kingdom
- ⁴ HIITScience, Revelstoke, BC, Canada
- ⁵ Institute for Health and Sport, Victoria University, Melbourne, VIC, Australia
- ⁶ Kitman Labs, Performance Research Intelligence Initiative, Dublin, Ireland
- 14
- 15 *Contact details:*
- 16 Mathieu Lacome
- 17 Paris Saint-Germain FC, Saint-Germain-en-Laye, France
- 18 Email: mlacome@psg.fr
- 19 Mobile: +33 6 09 42 78 33
- 20 Running head: Glucose monitoring in athletes
- 21 Abstract count: 217
- 22 Text only word account: 1900
- 23 Numbers of tables: 2
- 24 Numbers of figures: 3
- 25 Conflicts of interest: The authors do not have any conflict of interest.
- 26
- 27
- 28

29 Abstract

Purpose: To assess the concurrent validity of a continuous blood glucose monitoring system 30 (CGM) Post-Breakfast, Pre-exercise, Exercise and Post-exercise, while assessing the impact 31 of two different breakfasts on the observed level of validity. Methods: Eight non-diabetic 32 33 recreational athletes (age: 30.8±9.5 years; height: 173.6±6.6 cm; body mass: 70.3±8.1 kg) took part in the study. Blood glucose concentration was monitored every 10 min using both a 34 35 CGM (FreeStyle Libre, Abbott, France) and finger-prick blood glucose measurements (FreeStyle Optimum, Abbott, France) over 4 different periods (Post-Breakfast, Pre-Exercise, 36 37 Exercise and Post-Exercise). Two different breakfasts (carbohydrates- [CHO] and protein-[PROT] oriented) over two days (2x2 days in total) were used. Statistical analyses included 38 39 the Bland-Altman method, standardized mean bias (expressed in standardized unit), median absolute relative difference (MARD) and the Clarke Error Grid (EGA). Results: Overall, 40 41 mean bias was trivial-to-small at Post-Breakfast (effect size ± 90% confidence limits: -42 0.12±0.08), Pre-Exercise (-0.08±0.08) and Post-Exercise (0.25±0.14), while moderate during Exercise (0.66±0.09). Higher MARD was observed during Exercise (13.6% vs 7 to 9.5% for 43 the other conditions). While there was no effect of the breakfast type on the MARD results, 44 EGA revealed higher value in Zone D (i.e. clinically unsafe zone) during Exercise for CHO 45 (10.5%) compared with PROT (1.6%). Conclusion: The CGM device examined in this study 46 can only be validly used at rest, after both a CHO and PROT-rich breakfast. Using CGM to 47 monitor blood glucose concentration during exercise is not recommended. Moreover, the 48 accuracy decreased when carbohydrates are consumed before exercise. 49

50

51

52

53

54

55

56

57

58

59 Introduction

Regulation of blood glucose has first been widely studied from a health perspective. Hyperglycemia for example, is believed to be an independent risk factor for the development of several diseases such as type II diabetes mellitus¹ and cardiovascular disease.² More recently, the monitoring of blood glucose concentration has also elicited great interest in sport, as hypoglycaemia influences both physical and cognitive performances.³

In particular, it is known that at the beginning of exercise or after half-time in team sports, athletes experience transient hypoglycemia, which may affect physical and cognitive performance.⁴ Moreover, it has then been shown that a large glycemic variability exists among individuals in the general population.⁵ Additionally, similar results have been shown in sub-elite athletes,⁶ suggesting that providing more individualized guidelines to regulate blood glucose would be beneficial for both health and performance goals.

The emergence of new technologies such as continuous glucose monitoring (CGM) devices 71 has allowed blood glucose concentration dynamics to be captured more frequently and less 72 73 invasively than traditional measures such as finger pricks. Indeed, as CGM devices only need 74 to be placed once (usually on the back of the arm), it can be used for several days without disturbing sport practices. So far, these devices have been mainly used by diabetic populations 75 76 but as the technology becomes more accurate, less invasive, and less expensive, their use has increased in other populations and especially in healthy individuals. Therefore, the inclusion 77 78 of CGM in sport nutritionists' monitoring tool box could help to optimize nutritional strategies before and during exercise, and in turn, improve athletes' performance by 79 80 preventing hypoglycemia. However, to date, the validity of these new systems at rest or during exercise has been only assessed in diabetics patients and showed promising results.⁷ 81 82 Evidence regarding its relevance with an athletic population is still lacking. Moreover, the ability of such devices to detect potential glucose fluctuations due to different nutritional 83 intakes need to be confirmed. 84

Therefore, the aim of this study was to assess the concurrent validity of a new CGM device during different periods, *i.e.* pre, during and after exercise, while assessing the potential impact of different nutritional intakes in the observed level of validity.

88

89 Methodology

90 Study Population

Eight non-diabetic recreational athletes (5 females, 3 males) (age: 30.8 ± 9.5 years; height: 91 173.6 ± 6.6 cm; body mass: 70.3 ± 8.1 kg) who regularly participate in running and 92 resistance-based training (8±2 hours per week) were included in the study. An a priori power 93 94 analysis was conducted using the package pwr from R software (Version 4.0.0) for t-tests for non-parametric data with a significance level alpha of 0.05 a power of 0.8 and add a non-95 parametric correction of 15%. Result showed a minimal sample of 310 paired observations for 96 8 participants were necessary. Alcohol intake was prohibited during the study period. 97 98 Regarding female participants, we ensured they were all within the same menstrual phase during the study period. 99

Participants provided informed consent prior to starting the study. Ethics approval was
 granted before any data collection wwas undertaken and the recommendations of the
 Declaration of Helsinki were respected.

103

104

105 Design

A concurrent validity design was employed to assess the validity of a CGM system against 106 107 finger prick measures which was considered as the reference method. Over 2 consecutive weeks, participants took part in 4 nonconsecutive standardized days. Each standardized day 108 was broken-down into 4 distinct periods: 1) Post-Breakfast which corresponded to the first 109 110 hour after the end of the Breakfast 2) Pre-Exercise which corresponded to the first hour following the Post-Breakfast, 3) Exercise, which started 2 hours after the end of the breakfast 111 and lasted from the beginning of the warm up to the end of the workout and 4) Post-exercise, 112 which started immediately at the end of the workout, and up to 30 min later. A detailed 113 outline of the standardized day structure is provided in Figure 1. Nutritional intake during 114 breakfast was manipulated in order to provide either a high carbohydrate (CHO) or protein 115 (PROT) breakfast, to induce different levels of resting ad pre-exercise glycemia. Each typical 116 breakfast was repeated twice. Over those standardized days, blood glucose was measured 117 continuously with a CGM, while finger prick measures were taken every 10 minutes and. Day 118 119 1 was used for each participant to familiarize with the CGM and ensure calibration (as per 120 manufacturer recommendations) before the experimentation could start. Between day 2 and

121 13, participants undertook at their convenience the 4 standardized days. They were also122 instructed to have at least one full day of recovery between each experimental day.

123

124

Insert Figure 1

125

126 Methodology

127 *Continuous glucose monitoring.* Each participant was provided with a CGM system 128 (FreeStyle Libre, Abbott, France) over the full duration of the study. Each participant inserted 129 a sensor (FreeStyle Libre, Abbott, France) in their non-dominant upper arm (*i.e.* back the 130 triceps brachialis) one day before the beginning of the study. Glucose concentration was 131 recorded in the interstitial fluid every minute.

Finger prick blood glucose. Finger prick (FreeStyle Optium, Abbott, France) measures were collected following the procedure described by Gomez.⁸ Each sample was immediately analysed using the FreeStyle Libre reader (FreeStyle Libre Reader, Abbott, France) (The validity and reliability of this device has been previously confirmed.⁹

Breakfast. Two typical breakfasts were employed. The CHO breakfast contained a high 136 proportion of carbohydrates (CHO) with 1 gKg⁻¹ of body mass with a ceiling set at 70g of 137 carbohydrates per breakfast (e.g. breakfast contained a mix of orange juice, bread and 138 jam). The macronutrients and energy were as follow: 65±7g of carbohydrates, 9±1g of 139 proteins and 1±0g of fat for a total of 311±31 Kcal. The protein (PROT) breakfast was 140 isoenergetic compared with CHO (e.g. breakfast contained a mix of eggs, ham and cheese). 141 The macronutrients and energy were as follow: 1±0g of carbohydrates, 30±0g of proteins and 142 23±0g of fat for a total of 311±31 Kcal. 143

Standardized exercise. Participants completed the 30-15 Intermittent Fitness Test ($30-15_{IFT}$) as described by Buchheit et al.¹⁰ prior the beginning of the study. The speed (km·hr⁻¹) achieved by each participant during the last successfully completed stage of the test was recorded (V_{IFT}) in order to prescribe exercise intensity. The standardized exercise started with a 10-min low-intensity run (30 to 40% of V_{IFT}) and was followed by a high-intensity intermittent training exercise performed outdoor. The trials consisted of six reps of 3-min running intervals interspersed with 2 min of passive recovery. Reps 1 and 2 were performed at 151 75% V_{IFT} , reps 3 and 4 at 80% V_{IFT} and reps 5 and 6 at 85% V_{IFT} . The session was ended with 152 a 10-min walk.

Data processing. Each time point within a specific period was averaged as described above to perform the concurrent validity analysis for each method (CGM and finger prick) and per specific period (Figure 1). Each standardized day was analyzed first without (overall) and then as a function of breakfast type (CHO and PROT).

157 Statistical Analysis

Bland-Altmann method for repeated measures and standardized mean bias were first applied
to assess the agreement between CGM and finger prick measures at each specific period.¹¹
The following thresholds were applied to rate the magnitude of the bias as follow: >0.2
(small), >0.6 (moderate), >1.2 (large) and >2 (very large).¹²

Additionally, analysis of the median average relative difference (MARD)¹³ and the Clarke 162 Error Grid Analysis (EGA)¹⁴ were conducted. Regarding MARD, further comparisons 163 between the different periods were performed using Wilcoxon test and/or Kruskal-Wallis 164 tests. Level of statistical significance was set at P<0.05. Results were further analyzed while 165 166 calculating standardized differences, i.e. Wilcoxon effect sizes. The thresholds to rate the 167 magnitude of the effects were the same than those used for mean bias. Regarding EGA, results were divided into 5 zones (A, B, C, D, E). Each zone denotes a degree of clinical 168 implications of blood glucose concentration measures. Zones A and B were considered 169 clinically acceptable while zone C, D and E (erroneous treatment) were deemed possibly 170 unsafe.14 171

172

173 **Results**

The Bland-Altman analysis for the 4 periods is presented in Figure 2 and reported as mean 174 bias (standard error). Irrespectively of the breakfast content, mean biases were trivial-to-small 175 for Post-Breakfast (-2.99 [17.75] mg/dL), Pre-Exercise (-1.67 [10.95] mg/dL), Post-Exercise 176 (4.18 [17.88] mg/dL) and moderate during Exercise (12.25 [13.86] mg/dL). Regarding CHO 177 178 breakfast, mean biases were trivial-to-small for Post-Breakfast (-1.43 [25.98] mg/dL), Pre-Exercise (-4.29 [11.66] mg/dL), Post-Exercise (3.32 [18.18] mg/dL) and moderate during 179 Exercise (14.06 [13.81] mg/dL). For PROT Breakfast, trivial mean bias was observed for Pre-180 Exercise (0.91 [8.98] mg/dL), Post-Breakfast (-4.51 [8.31] mg/dL) and Post-Exercise (5.13 181

182 183	[15.98] mg/dL), while moderate mean biases were observed for Exercise (10.47 [13.19] mg/dL).
184	
185	**Insert Figure 2**
186	**Insert Figure 3**
187 188	The results of the MARD analysis between the different periods are presented in Table 1 and 2.
189	
190	**Insert Table 1 and 2**
191	
192 193 194 195 196 197 198 199	Results regarding EGA are presented in Table 3. Irrespectively of the breakfast content, Post-Breakfast, Pre-Exercise, and Post-Exercise periods fell into Zone A (accurate) and B (benign errors) (100%). However, during Exercise, 94% of the values fell into A (70.4%) and B (23.6%), and 6% in Zone D (failure to treat errors). For CHO breakfast, 10.5% of data fell into Zone D for Exercise, while the other periods fell into Zone A and B. Similarly, for PROT breakfast, 1.6% fell into Zone D during the Exercise period.
200	Discussion
201 202 203 204 205 206	The aim of this study was 1) to investigate the concurrent validity of a new CGM device in recreational athletes at Post-Breakfast, Pre-exercise, Exercise and Post-exercise, and 2) to assess the potential impact of either a CHO-rich or protein-rich breakfast on the observed level of validity. The main results highlighted that, while the validity of CGM was acceptable at rest (<i>i.e.</i> Post-Breakfast, Pre-Exercise and Post-Exercise), it was lower during Exercise and especially after the CHO breakfast.
207	The first results demonstrated trivial-to-small mean bias during all the non-exercise periods,
208 209	irrespectively of nutritional intake. Moreover, all results from EGA fell into the "clinically safe zone" (A and B), albeit during Exercise. These results are similar to those shown
205	sure zone (11 and D), about during Exercise. These results are similar to those shown

previously in non-athletic diabetic populations.¹⁵ Indeed, the present results suggest that

210

assessing glucose dynamics at rest is feasible with this CGM device. This could open the door
to a better individualization of nutritional strategies.⁵

Yet, we observed a higher bias during Exercise compared with the other periods, confirming 213 previous studies in a non-athletic diabetic population.¹⁶ Reasons that may contribute to the 214 215 reduced validity of the CGM device in this context include microcirculation perturbations as a as a result of movements around or within the insertion area, increases in body temperature 216 and rapid fluxes in glucose levels during exercise.¹⁷ Regarding the likely physiological time 217 lag of glucose transport between blood and interstitial fluid compartments (see Figure 3, 218 219 finger pricks measures changed faster Post-Breakfast than that of the CGM device), it should 220 be noted that it might not have accounted for the observed difference in accuracy as the 221 pattern is not only delayed but it varies with time and conditions. Indeed, while a clear hypoglycemia was observed with finger prick measures immediately at the start of exercise 222 223 (which was the expected physiological response), the CGM showed an increased blood 224 glucose response (Figure 2). Nonetheless, this discrepancy indicates that the CGM device was unable to detect a potential hypoglycemia observed at the onset of exercise, and could 225 therefore not be used to assess strategies aiming at preventing this phenomenon in practice. It 226 is worth mentioning that a trend for a better agreement was observed toward the end of the 227 exercise periods (Figure 2). If the duration of the exercise also affects the accuracy of CGM, it 228 means that while the device may not be suitable for sport including short and intermittent 229 exercise durations, its use could perhaps be considered during longer event such as cycling, 230 231 trail or triathlon. This potential better accuracy toward longer exercise duration highlights the need to conduct further research involving 1) longer exercise duration, 2) nutritional intake 232 during long endurance race 3) various exercise modalities and 4) different intensities. 233

234 To examine the potential effect of the absolute levels of glycemia on the validity of the CGM device, different breakfasts were proposed (CHO and PRO). Similar MARD and EGA results 235 236 were observed, suggesting that the CGM validity was not affected by the breakfast content during non-exercise periods (i.e. Post-Breakfast, Pre-Exercise, Post-Exercise). Specific pre-237 competition nutritional strategies can have a positive influence on both the acute running 238 performance among rugby league players¹⁸ or endurance athletes,¹⁹ and the chronic training 239 adaptations to training.²⁰ Consequently, the use of this CGM device could be considered by 240 practitioners willing to monitor glycemic responses before and after competition or training, 241 to ensure the efficacy of the nutritional strategies employed. 242

However, during the Exercise period, the CGM accuracy was modulated by breakfast content. 243 Indeed, a 10 times higher value in Zone D of the EGA (i.e. clinically unsafe) was observed 244 post CHO (10.5%) compared with post PROT (1.6%) breakfast. In our study, zone D 245 corresponds to the situation where finger prick measures indicate an hypoglycemic state 246 whereas CGM measures are within the normal range¹⁴ suggesting that CGM failed to detect 247 the hypoglycemia occurring during exercise after the CHO-rich breakfast. It is well known 248 there is a rapid drop of blood glucose concentration at the onset of exercise, due to an 249 increased glucose uptake by exercising muscles.²¹ This physiological mechanism could 250 explain why the sensor lacks sensitivity to rapid changes in glucose concentration, as 251 observed in the present study. As it stands, if practitioners want to monitor blood glucose 252 during high-intensity intermittent exercise, they need to consider other devices than CGM 253 (e.g. finger prick). 254

255

256 **Practical applications**

257

The present CGM system provided valid measures at rest. Therefore, the use of such a
 system may allow for a better individualization of nutritional strategies before or after
 competition.

261

The level of validity was lower during high-intensity intermittent training and was in
 addition influenced by the type of breakfast consumed (i.e. high carbohydrates or high
 protein). Consequently, practitioners should avoid using this device during intermittent
 exercise.

266

267 Conclusion

Daily monitoring of blood glucose is of importance in athletes given the likely impact of glycemia on performance and the individualized nutritional recommendations that can be made with CGM. Our results highlighted that the CGM device examined in the present study presented only trivial-to-small bias when compared with a traditional fingerpick device at rest, suggesting that it could be used confidently during this specific period. The CGM device is not valid enough to monitor glucose during intermittent exercise. Further analyses should however evaluate the validity of this device over longer exercise duration.

275

Reference

277 278	1.	Cavalot F, Pagliarino A, Valle M, et al. Postprandial blood glucose predicts cardiovascular events and all-cause mortality in type 2 diabetes in a 14-year follow-up:
279		Lessons from the San Luigi Gonzaga diabetes study. <i>Diabetes Care</i> .
280		2011;34(10):2237-2243. doi:10.2337/dc10-2414
281	2.	Gallwitz B. Implications of postprandial glucose and weight control in people with type
282		2 diabetes: understanding and implementing the International Diabetes Federation
283		guidelines. Diabetes Care. 2009;32 Suppl 2(suppl 2):S322-S325. doi:10.2337/dc09-
284		s331
285	3.	Harper LD, Briggs MA, McNamee G, et al. Physiological and performance effects of
286		carbohydrate gels consumed prior to the extra-time period of prolonged simulated
287		soccer match-play. J Sci Med Sport. 2016;19(6):509-514.
288		doi:10.1016/j.jsams.2015.06.009
289	4.	Kingsley M, Penas-Ruiz C, Terry C, Russell M. Effects of carbohydrate-hydration
290		strategies on glucose metabolism, sprint performance and hydration during a soccer
291		match simulation in recreational players. J Sci Med Sport. 2014;17(2):239-243.
292		doi:10.1016/j.jsams.2013.04.010
293	5.	Hall H, Perelman D, Breschi A, et al. Glucotypes reveal new patterns of glucose
294		dysregulation. PLoS Biol. 2018;16(7):1-23. doi:10.1371/journal.pbio.2005143
295	6.	Thomas F, Pretty CG, Desaive T, Chase JG. Blood Glucose Levels of Subelite Athletes
296		during 6 Days of Free Living. J Diabetes Sci Technol. 2016;10(6):1335-1343.
297		doi:10.1177/1932296816648344
298	7.	Greene J, Louis J, Korostynska O, Mason A. State-of-the-art methods for skeletal
299		muscle glycogen analysis in athletes-the need for novel non-invasive techniques.
300		Biosensors. 2017;7(1):1-16. doi:10.3390/bios7010011
301	8.	Gómez AM, Umpierrez GE, Muñoz OM, et al. Continuous Glucose Monitoring Versus
302		Capillary Point-of-Care Testing for Inpatient Glycemic Control in Type 2 Diabetes
303		Patients Hospitalized in the General Ward and Treated with a Basal Bolus Insulin
304		Regimen. J Diabetes Sci Technol. 2016;10(2):325-329.
305		doi:10.1177/1932296815602905

306 307 308	9.	Rodrigo EP, Deib-Morgan K, Diego OG de, García-Velasco P, Sgaramella GA, González IG. Accuracy and reliability between glucose meters: A study under normal clinical practice conditions. <i>Semer - Med Fam.</i> 2017;43(1):20-27.
309 310 311	10.	Buchheit M. The 30-15 Intermittent Fitness Test: Accuracy for Individualizing Interval Training of Young Intermittent Sport Players. <i>J Strength Cond Res.</i> 2008;22(2):365-374.
312	11.	Altman D, Bland J. sensitivity.pdf. BMJ. 1994;308:1552.
313 314	12.	Hopkins WG. Measures of Reliability in Sports Medicine and Science. <i>Sport Med</i> . 2000;30(1):1-15.
315 316 317	13.	Reiterer F, Polterauer P, Schoemaker M, et al. Significance and Reliability of MARD for the Accuracy of CGM Systems. <i>J Diabetes Sci Technol</i> . 2017;11(1):59-67. doi:10.1177/1932296816662047
318 319 320	14.	Clarke WL, Cox D, Gonder-Frederick LA, Carter W, Pohl SL. Evaluating clinical accuracy of systems for self-monitoring of blood glucose. <i>Diabetes Care</i> . 1987;10(5):622-628. doi:10.2337/diacare.10.5.622
321 322 323 324	15.	Freckmann G, Pleus S, Link M, Zschornack E, Kloïtzer HM, Haug C. Performance evaluation of three continuous glucose monitoring systems: Comparison of six sensors per subject in parallel. <i>J Diabetes Sci Technol</i> . 2013;7(4):842-853. doi:10.1177/193229681300700406
325 326 327	16.	Biagi L, Bertachi A, Quirós C, et al. Accuracy of continuous glucose monitoring before, during, and after aerobic and anaerobic exercise in patients with type 1 diabetes mellitus. <i>Biosensors</i> . 2018;8(1):1-8. doi:10.3390/bios8010022
328 329 330	17.	Kumareswaran K, Elleri D, Allen JM, et al. Accuracy of continuous glucose monitoring during exercise in type 1 diabetes pregnancy. <i>Diabetes Technol Ther</i> . 2013;15(3):223-229. doi:10.1089/dia.2012.0292
331 332 333	18.	Bradley WJ, Morehen JC, Haigh J, et al. Muscle glycogen utilisation during Rugby match play: Effects of pre-game carbohydrate. <i>J Sci Med Sport</i> . 2016;19(12):1033-1038. doi:10.1016/j.jsams.2016.03.008
334 335	19.	Rothschild JA, Kilding AE, Plews DJ. What should i eat before exercise? Pre-exercise nutrition and the response to endurance exercise: Current prospective and future

336		directions. Nutrients. 2020;12(11):1-23. doi:10.3390/nu12113473
337 338	20.	Jeukendrup AE. Periodized Nutrition for Athletes. <i>Sport Med.</i> 2017;47(s1):51-63. doi:10.1007/s40279-017-0694-2
339 340	21.	Richter EA, Hargreaves M. Exercise, GLUT4, and skeletal muscle glucose uptake. <i>Physiol Rev.</i> 2013;93(3):993-1017. doi:10.1152/physrev.00038.2012
341		
342		

Table and figure caption

Figure 1. Schematic representation of the study design.

Figure 2. Bland-Altman analysis between the continuous glucose monitoring device (CGM) and finger prick measures (FPBG). Dash lines represent the limits of agreements.

Figure 3. Continuous glucose monitoring (CGM) and finger prick measures during each standardized condition, when ingesting a carbohydrate- (upper) and protein- (lower) oriented breakfasts, with the 2 days of each breakfast condition pooled for each participant ($n = 2 \ge 8$ for each curve). Data are presented as mean (SE).


Table 1. Median Absolute Relative Difference between the continuous glucose monitoring device (CGM) and finger prick measures. Data are median (interquartile range) and expressed in percentage. *: significantly different from Post-Breakfast. #: significantly different from Pre-Exercise. [†]: significantly different from Exercise. Comparisons between period are presented as effect size with 90% confidence interval.

Table 2. Comparisons between period are presented as effect size for Wilcoxon test with 90% confidence interval.


Table 3. Clark Error Grid Analysis between the continuous glucose monitoring device (CGM) and finger prick measures. Zone A represents a clinically accurate measure. Zone B stands for benign errors. Zone C represents overcorrection errors. Zone D and E represent failure to treat errors and erroneous treatment errors respectively. For more details see Clarke et al. (1987).

Figures


Figure 1














Tables

Table 1

	Post- Breakfast	Pre- Exercise	Exercise	Post- Exercise
Overall	9.1	7.1	13.6	9.4
	(4.6-13.8)	(3.6-13.4) [#]	(6.8-23.2) [*]	(5.0-17.3) ^{#†}
СНО	9.4	7.1	16.2	10.1
	(5.3-16.8)	(3.9-13.2) [*]	(7.4-25.6) ^{*#}	(6.1-16.9) ^{#†}
PROT	8.8	7.0	11.3	8.2
	(4-11.9)	(3.4-13.4)	(6-19.7) ^{*#}	(4.1-17.3)

	Post-Breakfast	Post-Breakfast	Pre-Exercise	Pre-Exercise	Exercise
	vs.	vs.	vs.	vs.	vs.
	Exercise	Post-Exercise	Exercise	Post-Exercise	Post Exercise
Overall	0.24	0.07	0.31	0.16	0.15
	(0.17 to 0.31)	(0.01 to 0.16)	(0.24 to 0.38)	(0.07 to 0.24)	(0.06 to 0.23)
СНО	0.24	0.06	0.37	0.19	0.18
	(0.13 to 0.34)	(0.01 to 0.18)	(0.27 to 0.46)	(0.07 to 0.31)	(0.07 to 0.28)
PROT	0.24	0.08	0.26	0.18	0.12
	(0.14 to 0.34)	(0.01 to 0.2)	(0.16 to 0.36)	(0.01 to 0.24)	(0.02 to 0.24)

Zone	Post-Breakfast	Pre-Exercise	Exercise	Post-Exercise

	A (Accurate)	189 (88.3%)	213 (93.4%)	176 (70.4%)	100 (76.3%)	Table 3
Overall	B (Benign errors)	25 (11.7%)	14 (6.1%)	59 (23.6%)	31 (23.7%)	
	D (Failure to treat errors)	/	1 (0.5%)	15 (6.0%)	/	
	A (Accurate)	85 (80.2%)	104 (92.0%)	81 (65.3%)	52 (75.4%)	
СНО	B (Benign errors)	21 (19.8%)	9 (8.0%)	30 (24.2%)	17 (24.7%)	
	D (Failure to treat errors)	/	/	13 (10.5%)	/	
	A (Accurate)	104 (96.3%)	109 (94.8%)	95 (75.4%)	48 (77.4%)	
PROT	B (Benign errors)	4 (3.7%)	5 (4.3%)	29 (23.0%)	14 (22.6%)	
	D (Failure to treat errors)	/	1 (0.9%)	2 (1.6%)	/	